

Vitalograph micro

Name _____ ID _____

Age: 25 Sex: Male Height: 170cm Weight: 80 kg
Predicted Values Set: ERS Polgar Correction: 100%
Accuracy Check: 27/01/2006 Calibration Adjusted: 27/01/2006
Micro Serial #: 03692


Test Information:

Test Date: 01/01/2005 Test Time: 13:38
Number Of Tests: 1 ATS/ERS Best : 1
FVC Rep.: - FEV1 Rep.: - PEF Rep.: -


Values at B.T.P.S

| Index | | Pred. | Meas. Best | % Pred. |
|----------|-------|-------|---------------|---------|
| FVC | L | 5.04 | 3.45 | 68 |
| FEV1 | L | 4.19 | 2.76 | 66 |
| FEV1% | % | 83 | 80 | 96 |
| FEV6 | L | 5.02 | - | - |
| PEF | L/min | 571 | 318 | 56 |
| FEF25 | L/s | 8.09 | 5.29 | 65 |
| FEF75 | L/s | 2.45 | 1.21 | 49 |
| FEF25-75 | L/s | 4.44 | 2.55 | 57 |

F(L/S)



V(L)



Comments:

Computer Suggested Interpretation: Mild Restriction

Signature: _____

Software Reference: 63130 v2.00